
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000750
9:05 27/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:14
STERILIZING 4:24
136.4 C 232kPa 4:24
136.2 C 231kPa 4:54
136.1 C 231kPa 5:24
136.3 C 232kPa 5:54
136.4 C 232kPa 6:24
136.1 C 230kPa 6:54
136.4 C 233kPa 7:24
136.4 C 233kPa 7:54
136.4 C 233kPa 7:54
VENTING 7:54
AIR DRYING 8:48
CYCLE COMPLETE 8:50

Digital Signature #
E1F8C335F5708292
