
Midmark M11 - v1.0.0
Total Cycles: 5123

Name: RITTER 1
Sterilizer ID: V2178198
11/21/2024 04:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.2 F 0.0
2:00 146.5 F 0.0
4:00 166.2 F 0.5
6:00 191.5 F 2.2
8:00 210.3 F 6.0
10:00 222.5 F 10.4
12:00 229.2 F 7.4
14:00 238.6 F 11.5
16:00 236.3 F 9.3
18:00 238.4 F 9.5
20:00 248.7 F 14.7
22:00 260.6 F 21.2
24:00 271.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.4 F 29.1
2:00 271.8 F 28.5
3:00 271.3 F 28.5
4:00 272.4 F 29.1

Min 271.2 F 28.2
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 212.8 F 0.0
10:00 228.1 F 0.0
15:00 238.0 F 0.0
20:00 225.5 F 0.0
25:00 221.7 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:44
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:03:25
CYCLE COMPLETE

11/21/2024 05:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____