
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000220
7:05 09/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:52
STERILIZING 3:50
136.0 C 220kPa 3:50
136.1 C 220kPa 4:20
136.2 C 222kPa 4:50
136.2 C 220kPa 5:20
136.4 C 223kPa 5:50
136.5 C 223kPa 6:20
136.4 C 223kPa 6:50
136.2 C 219kPa 7:20
136.2 C 219kPa 7:20
VENTING 7:20
AIR DRYING 8:09
CYCLE COMPLETE 8:43

Digital Signature #
8678E844E734D59A
