

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000610  
11:32 27/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:13  
PRESSURIZING 2:19  
STERILIZING 3:26  
136.5 C 220kPa 3:26  
136.2 C 231kPa 3:56  
136.4 C 232kPa 4:26  
136.4 C 233kPa 4:56  
136.3 C 232kPa 5:26  
136.3 C 232kPa 5:56  
136.4 C 230kPa 6:26  
136.4 C 233kPa 6:56  
136.4 C 233kPa 6:56  
VENTING 6:56  
AIR DRYING 7:47  
CASSETTE REMOVED  
CYCLE COMPLETE 8:48  
-----

Digital Signature #  
6818993971F61241  
-----