
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001622
10:15 27/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:37
STERILIZING 3:38
136.0 C 220kPa 3:38
136.4 C 223kPa 4:08
136.4 C 225kPa 4:38
136.0 C 221kPa 5:08
136.2 C 223kPa 5:38
136.3 C 223kPa 6:08
136.3 C 223kPa 6:38
136.1 C 223kPa 7:08
136.1 C 223kPa 7:08
VENTING 7:08
AIR DRYING 7:57
CYCLE COMPLETE 8:02

Digital Signature #
D385631A99352513
