
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000387
14:00 01/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:38
STERILIZING 4:35
136.0 C 220kPa 4:35
135.8 C 221kPa 5:05
136.4 C 224kPa 5:35
136.2 C 223kPa 6:05
136.4 C 225kPa 6:35
136.4 C 225kPa 7:05
136.4 C 225kPa 7:35
136.3 C 224kPa 8:05
136.3 C 224kPa 8:05
VENTING 8:05
AIR DRYING 9:00
CYCLE COMPLETE 9:06

Digital Signature #
B34ECBBA90DFECA6
