

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001062  
11:10 30/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:06  
PRESSURIZING 3:09  
STERILIZING 4:14  
136.0 C 220kPa 4:14  
136.4 C 221kPa 4:44  
136.4 C 225kPa 5:14  
136.4 C 223kPa 5:44  
136.4 C 223kPa 6:14  
136.2 C 222kPa 6:44  
136.4 C 224kPa 7:14  
136.1 C 222kPa 7:44  
136.1 C 222kPa 7:44  
VENTING 7:44  
AIR DRYING 8:36  
CYCLE COMPLETE 8:43  
-----

Digital Signature #  
149BB2E7D26FB08D  
-----