
Midmark M11 - v1.0.0
Total Cycles: 5048

Name: RITTER 1
Sterilizer ID: V2178198
11/04/2024 09:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 75.1 F 0.0
4:00 77.8 F 0.0
6:00 108.3 F 0.1
8:00 139.8 F 0.3
10:00 170.7 F 0.8
12:00 192.4 F 3.9
14:00 209.3 F 7.0
16:00 219.6 F 10.6
18:00 223.7 F 7.5
20:00 231.7 F 8.9
22:00 242.6 F 13.6
24:00 241.0 F 10.9
26:00 237.2 F 8.9
28:00 251.2 F 15.5
30:00 262.4 F 21.9
32:00 272.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.4 F 28.3
2:00 271.9 F 28.3
3:00 272.2 F 28.2
4:00 271.9 F 28.3

Min 271.5 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.1 F 0.0
5:00 200.1 F 0.0
10:00 196.8 F 0.0
15:00 209.6 F 0.0
20:00 223.0 F 0.0
25:00 233.1 F 0.0
30:00 231.5 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 32:22
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:10:32
CYCLE COMPLETE

11/04/2024 10:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____