
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000785
13:03 10/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:40
PRESSURIZING 6:31
STERILIZING 8:03
136.2 C 220kPa 8:03
136.1 C 218kPa 8:33
135.9 C 219kPa 9:03
136.4 C 223kPa 9:33
136.4 C 223kPa 10:03
136.1 C 221kPa 10:33
136.3 C 221kPa 11:03
136.4 C 221kPa 11:33
136.4 C 221kPa 11:33
VENTING 11:33
AIR DRYING 12:33
CYCLE COMPLETE 23:47

Digital Signature #
F53D3BC2B2C425D0
