

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4745

Name: RITTER 1  
Sterilizer ID: V2178198  
07/18/2024 01:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 88.1 F 0.0  
2:00 90.1 F 0.5  
4:00 101.0 F 2.0  
6:00 120.8 F 4.4  
8:00 148.8 F 7.1  
10:00 187.9 F 10.2  
12:00 204.1 F 10.2  
14:00 211.5 F 5.7  
16:00 221.5 F 9.2  
18:00 231.4 F 13.1  
20:00 231.6 F 8.7  
22:00 241.8 F 13.5  
24:00 240.7 F 11.1  
26:00 251.7 F 16.6  
28:00 262.3 F 22.5  
30:00 272.0 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 271.5 F 29.0  
2:00 272.8 F 29.8  
3:00 272.3 F 29.3  
4:00 271.5 F 29.0

Min 271.0 F 28.5  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.1 F 0.0  
5:00 209.6 F 0.0  
10:00 199.8 F 0.0  
15:00 203.2 F 0.0  
20:00 210.9 F 0.0  
25:00 221.2 F 0.0  
30:00 234.0 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 30:26  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:09:05  
CYCLE COMPLETE

07/18/2024 02:10 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_