
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000662
14:28 13/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:41
STERILIZING 3:47
136.0 C 230kPa 3:47
136.2 C 230kPa 4:17
136.3 C 231kPa 4:47
136.2 C 231kPa 5:17
136.4 C 233kPa 5:47
136.4 C 229kPa 6:17
136.4 C 233kPa 6:47
136.0 C 229kPa 7:17
136.0 C 229kPa 7:17
VENTING 7:17
AIR DRYING 8:13
CYCLE COMPLETE 8:18

Digital Signature #
4D5CA54D096630AA
