

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001542  
11:12 24/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:09  
STERILIZING 4:07  
136.3 C 220kPa 4:07  
136.2 C 223kPa 4:37  
136.3 C 225kPa 5:07  
136.2 C 223kPa 5:37  
136.3 C 226kPa 6:07  
136.1 C 224kPa 6:37  
136.2 C 223kPa 7:07  
136.4 C 226kPa 7:37  
136.4 C 226kPa 7:37  
VENTING 7:37  
AIR DRYING 8:31  
CYCLE COMPLETE 11:02

-----  
Digital Signature #  
C1D18C5C7786602A  
-----