
Midmark M11 - v1.0.0
Total Cycles: 2560

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2022 11:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.4 F 0.0
2:00 191.0 F 4.0
4:00 198.7 F 9.4
6:00 215.8 F 5.9
8:00 231.2 F 9.8
10:00 239.0 F 8.5
12:00 243.5 F 13.3
14:00 243.6 F 12.2
16:00 256.9 F 19.2
18:00 268.5 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.7 F 29.0
2:00 271.7 F 28.5
3:00 272.8 F 29.1
4:00 271.9 F 28.4

Min 271.6 F 28.3
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.9 F 0.0
5:00 202.0 F 0.0
10:00 210.9 F 0.0
15:00 216.7 F 0.0
20:00 205.3 F 0.0
25:00 198.1 F 0.0
30:00 194.2 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

11/15/2022 12:56 PM
ACCEPTED: X
REJECTED:
Operator: U02