
Midmark M11 - v1.0.5
Total Cycles: 4308

Name: RITTER2
Sterilizer ID: V2313228
01/28/2025 07:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.0 F 0.0
2:00 127.5 F 1.4
4:00 136.5 F 4.5
6:00 172.0 F 8.1
8:00 188.9 F 12.4
10:00 203.9 F 5.8
12:00 218.3 F 10.0
14:00 225.3 F 9.2
16:00 236.1 F 11.7
18:00 235.6 F 9.9
20:00 248.6 F 16.0
22:00 259.6 F 22.3
24:00 270.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 30.0
1:00 272.7 F 30.9
2:00 272.3 F 30.4
3:00 271.5 F 29.9
4:00 271.8 F 30.5

Min 271.3 F 29.9
Max 273.1 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 207.6 F 0.0
10:00 207.7 F 0.0
15:00 204.2 F 0.0
20:00 195.8 F 0.0
25:00 197.9 F 0.0
30:00 202.9 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 25:01
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:03:57
CYCLE COMPLETE

01/28/2025 08:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____