
Midmark M11 - v1.0.0
Total Cycles: 1138

Name: RITTER 1
Sterilizer ID: V2178198
09/02/2021 06:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 80.2 F 0.0
4:00 86.4 F 0.0
6:00 129.8 F 0.0
8:00 163.3 F 0.1
10:00 188.5 F 0.3
12:00 207.7 F 2.7
14:00 221.7 F 5.7
16:00 229.3 F 8.9
18:00 238.4 F 12.6
20:00 235.2 F 8.7
22:00 241.9 F 11.1
24:00 244.9 F 12.5
26:00 242.7 F 11.4
28:00 256.2 F 18.3
30:00 267.7 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.5 F 28.4
2:00 272.3 F 28.5
3:00 272.0 F 28.0
4:00 272.7 F 28.8

Min 271.5 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 165.8 F 0.0
5:00 207.4 F 0.0
10:00 195.2 F 0.0
15:00 193.9 F 0.0
20:00 192.8 F 0.0
25:00 190.7 F 0.0
30:00 188.2 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 31:18
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:09:28
CYCLE COMPLETE

09/02/2021 08:28 AM
ACCEPTED: X
REJECTED:
Operator: U02