
Midmark M11 - v1.0.0
Total Cycles: 3770

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2023 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.2 F 0.0
2:00 162.8 F 1.6
4:00 180.9 F 4.9
6:00 202.9 F 9.4
8:00 215.4 F 8.9
10:00 227.9 F 8.1
12:00 238.1 F 13.1
14:00 239.7 F 10.3
16:00 236.4 F 8.4
18:00 249.0 F 14.5
20:00 260.2 F 20.7
22:00 270.3 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 28.7
2:00 272.4 F 29.1
3:00 272.1 F 28.5
4:00 272.6 F 29.1

Min 271.5 F 28.4
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.7 F 0.0
5:00 231.4 F 0.0
10:00 243.5 F 0.0
15:00 219.3 F 0.0
20:00 214.2 F 0.0
25:00 216.4 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 23:37
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:02:24
CYCLE COMPLETE

09/14/2023 10:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____