

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4185

Name: RITTER 1  
Sterilizer ID: V2178198  
01/10/2024 08:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.4 F 0.0  
2:00 74.8 F 0.3  
4:00 111.8 F 1.3  
6:00 115.1 F 3.7  
8:00 141.8 F 6.9  
10:00 168.3 F 10.8  
12:00 192.9 F 2.9  
14:00 209.8 F 5.1  
16:00 220.7 F 8.5  
18:00 231.9 F 12.4  
20:00 231.8 F 8.2  
22:00 243.8 F 12.8  
24:00 242.4 F 11.8  
26:00 255.4 F 18.4  
28:00 266.6 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.1 F 28.8  
2:00 272.8 F 29.3  
3:00 272.0 F 29.0  
4:00 272.0 F 28.7

Min 271.5 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 212.2 F 0.0  
10:00 226.3 F 0.0  
15:00 230.1 F 0.0  
20:00 223.5 F 0.0  
25:00 221.7 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 29:32  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:07:32  
CYCLE COMPLETE

01/10/2024 09:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_