
Midmark M11 - v1.0.5
Total Cycles: 2878

Name: RITTER2
Sterilizer ID: V2313228
09/13/2023 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.3 F 0.0
2:00 167.9 F 2.8
4:00 182.4 F 6.9
6:00 205.5 F 12.2
8:00 222.6 F 6.4
10:00 236.9 F 11.6
12:00 234.4 F 9.0
14:00 239.1 F 10.1
16:00 247.9 F 14.5
18:00 260.3 F 21.4
20:00 271.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.8
1:00 272.7 F 29.5
2:00 271.7 F 28.8
3:00 272.4 F 29.2
4:00 271.7 F 29.0

Min 271.6 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 228.8 F 0.0
10:00 238.7 F 0.0
15:00 229.1 F 0.0
20:00 225.5 F 0.0
25:00 223.0 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

09/13/2023 02:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____