
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000790
15:11 11/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:34
PRESSURIZING 7:36
STERILIZING 9:13
136.0 C 220kPa 9:13
136.2 C 222kPa 9:43
136.2 C 221kPa 10:13
136.4 C 223kPa 10:43
136.4 C 221kPa 11:13
136.4 C 224kPa 11:43
136.4 C 223kPa 12:13
136.3 C 223kPa 12:43
136.3 C 223kPa 12:43
VENTING 12:43
AIR DRYING 13:47
CYCLE COMPLETE 43:47

Digital Signature #
0A1E09B5F08CC96A
