
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008280
9:29 31/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:27
STERILIZING 3:21
136.0 C 220kPa 3:21
136.3 C 222kPa 3:51
136.2 C 220kPa 4:21
136.4 C 222kPa 4:51
136.4 C 223kPa 5:21
136.3 C 220kPa 5:51
136.4 C 221kPa 6:21
136.2 C 222kPa 6:51
136.2 C 222kPa 6:51
VENTING 6:51
AIR DRYING 7:42
CYCLE COMPLETE 37:42

Digital Signature #
20F45A89FA2EFD68
