
Midmark M11 - v1.0.0
Total Cycles: 5415

Name: RITTER 1
Sterilizer ID: V2178198
03/14/2025 08:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 95.4 F 0.3
4:00 87.2 F 1.5
6:00 123.4 F 3.8
8:00 154.7 F 6.7
10:00 208.4 F 10.5
12:00 193.0 F 4.5
14:00 209.8 F 5.1
16:00 222.1 F 8.5
18:00 231.4 F 12.3
20:00 230.4 F 7.1
22:00 241.5 F 11.9
24:00 237.1 F 9.0
26:00 249.8 F 15.0
28:00 260.6 F 21.1
30:00 270.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.9 F 28.9
2:00 271.5 F 28.1
3:00 272.5 F 29.0
4:00 272.2 F 28.4

Min 271.5 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 174.9 F 0.0
10:00 187.3 F 0.0
15:00 193.6 F 0.0
20:00 195.5 F 0.0
25:00 196.7 F 0.0
30:00 198.3 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 30:49
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:09:01
CYCLE COMPLETE

03/14/2025 09:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____