

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2911

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2023 01:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.5 F 0.0  
2:00 177.1 F 3.3  
4:00 188.6 F 7.5  
6:00 206.0 F 12.0  
8:00 217.6 F 6.5  
10:00 228.4 F 11.3  
12:00 232.4 F 8.6  
14:00 240.1 F 11.0  
16:00 245.0 F 13.5  
18:00 256.8 F 19.8  
20:00 267.4 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.1  
1:00 272.6 F 29.7  
2:00 271.8 F 29.5  
3:00 272.6 F 29.8  
4:00 272.7 F 30.1

Min 271.0 F 29.0  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.0 F 0.0  
5:00 220.7 F 0.0  
10:00 228.5 F 0.0  
15:00 209.7 F 0.0  
20:00 207.8 F 0.0  
25:00 208.2 F 0.0  
30:00 205.8 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 21:26  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:00:38  
CYCLE COMPLETE

09/20/2023 02:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_