
Midmark M11 - v1.0.5
Total Cycles: 1915

Name: RITTER2
Sterilizer ID: V2313228
01/10/2023 04:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.6 F 0.0
2:00 117.4 F 0.0
4:00 150.7 F 0.3
6:00 177.5 F 4.2
8:00 198.2 F 8.9
10:00 214.0 F 9.9
12:00 227.0 F 8.4
14:00 239.1 F 13.5
16:00 240.1 F 11.3
18:00 239.2 F 10.2
20:00 253.2 F 17.2
22:00 265.2 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.2
2:00 271.8 F 29.1
3:00 272.5 F 29.2
4:00 271.5 F 28.8

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 229.2 F 0.0
10:00 225.6 F 0.0
15:00 225.1 F 0.0
20:00 219.1 F 0.0
25:00 222.2 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 23:43
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:58
CYCLE COMPLETE

01/10/2023 05:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____