

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5445

Name: RITTER 1  
Sterilizer ID: V2178198  
03/25/2025 04:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.2 F 0.0  
2:00 175.8 F 3.7  
4:00 200.7 F 8.5  
6:00 218.3 F 7.3  
8:00 229.7 F 8.0  
10:00 241.9 F 13.0  
12:00 239.0 F 10.2  
14:00 236.1 F 8.1  
16:00 249.9 F 15.1  
18:00 261.9 F 21.8  
20:00 272.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 271.9 F 28.6  
2:00 271.6 F 28.6  
3:00 271.8 F 28.5  
4:00 271.6 F 28.7  
  
Min 271.2 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.8 F 0.0  
5:00 200.5 F 0.0  
10:00 197.5 F 0.0  
15:00 199.7 F 0.0  
20:00 199.2 F 0.0  
25:00 241.2 F 0.0  
30:00 231.2 F 0.0  
Drying Complete

FILLING: 1:41  
HEATING: 20:21  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 00:59:38  
CYCLE COMPLETE

03/25/2025 05:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_