
Midmark M11 - v1.0.0
Total Cycles: 3367

Name: RITTER 1
Sterilizer ID: V2178198
06/09/2023 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.3 F 0.0
2:00 140.2 F 0.0
4:00 169.4 F 1.2
6:00 190.2 F 5.3
8:00 207.8 F 9.7
10:00 218.4 F 8.5
12:00 228.2 F 8.4
14:00 239.4 F 13.2
16:00 239.1 F 10.6
18:00 236.1 F 8.7
20:00 250.5 F 15.5
22:00 262.0 F 22.0
24:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.8 F 28.9
2:00 271.6 F 28.2
3:00 272.8 F 29.1
4:00 272.4 F 28.6

Min 271.6 F 28.1
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 202.4 F 0.0
10:00 221.2 F 0.0
15:00 221.5 F 0.0
20:00 220.1 F 0.0
25:00 218.6 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:27
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:02:40
CYCLE COMPLETE

06/09/2023 01:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____