
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006078
14:20 04/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 2:58
STERILIZING 4:00
136.0 C 220kPa 4:00
136.3 C 222kPa 4:30
136.1 C 220kPa 5:00
136.2 C 223kPa 5:30
135.9 C 220kPa 6:00
136.3 C 224kPa 6:30
136.3 C 223kPa 7:00
136.0 C 221kPa 7:30
136.0 C 221kPa 7:30
VENTING 7:30
AIR DRYING 8:20
CYCLE COMPLETE 38:20

Digital Signature #
9721EA026AA24C61
