
Midmark M11 - v1.0.0
Total Cycles: 4846

Name: RITTER 1
Sterilizer ID: V2178198
08/23/2024 08:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 85.0 F 0.0
4:00 86.4 F 0.1
6:00 118.9 F 0.3
8:00 146.8 F 0.7
10:00 175.2 F 1.7
12:00 209.1 F 4.8
14:00 216.3 F 8.3
16:00 226.9 F 11.9
18:00 225.4 F 5.6
20:00 236.2 F 10.2
22:00 235.3 F 8.3
24:00 244.1 F 12.3
26:00 240.9 F 10.5
28:00 253.6 F 16.8
30:00 264.6 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.5 F 28.4
2:00 271.6 F 28.1
3:00 272.3 F 28.4
4:00 271.6 F 28.1

Min 271.4 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.8 F 0.0
5:00 226.1 F 0.0
10:00 225.3 F 0.0
15:00 230.8 F 0.0
20:00 220.1 F 0.0
25:00 223.0 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 0:40
HEATING: 31:58
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:57
CYCLE COMPLETE

08/23/2024 09:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____