
Midmark M11 - v1.0.0
Total Cycles: 3686

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2023 03:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.3 F 0.0
2:00 160.5 F 3.5
4:00 183.4 F 7.7
6:00 204.5 F 12.6
8:00 220.6 F 6.9
10:00 234.0 F 12.3
12:00 235.6 F 9.3
14:00 237.8 F 9.2
16:00 247.6 F 14.0
18:00 259.5 F 20.8
20:00 270.4 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.6 F 29.2
2:00 271.9 F 29.0
3:00 272.4 F 29.1
4:00 271.9 F 29.0

Min 271.5 F 28.6
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 230.4 F 0.0
10:00 226.1 F 0.0
15:00 221.9 F 0.0
20:00 212.9 F 0.0
25:00 230.7 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:28
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:06
CYCLE COMPLETE

08/25/2023 04:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____