

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4434

Name: RITTER2  
Sterilizer ID: V2313228  
03/19/2025 10:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 103.6 F 0.0  
2:00 103.2 F 0.5  
4:00 107.5 F 1.7  
6:00 137.3 F 4.5  
8:00 168.3 F 7.8  
10:00 193.4 F 11.5  
12:00 216.6 F 4.5  
14:00 229.6 F 8.6  
16:00 231.5 F 12.8  
18:00 233.9 F 9.0  
20:00 240.8 F 11.6  
22:00 242.7 F 12.2  
24:00 253.7 F 17.9  
26:00 263.9 F 24.0  
28:00 272.3 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 272.0 F 30.0  
2:00 272.2 F 30.2  
3:00 272.5 F 30.0  
4:00 271.9 F 29.9

Min 270.4 F 29.0  
Max 273.2 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.8 F 0.0  
5:00 227.1 F 0.0  
10:00 210.6 F 0.0  
15:00 210.5 F 0.0  
20:00 208.7 F 0.0  
25:00 225.7 F 0.0  
30:00 225.8 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 28:03  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:06:43  
CYCLE COMPLETE

03/19/2025 11:56 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_