
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000802
10:20 04/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:15
STERILIZING 4:23
136.0 C 228kPa 4:23
135.9 C 231kPa 4:53
136.4 C 233kPa 5:23
136.4 C 233kPa 5:53
136.0 C 231kPa 6:23
136.3 C 232kPa 6:53
136.3 C 231kPa 7:23
136.4 C 232kPa 7:53
136.4 C 232kPa 7:53
VENTING 7:53
AIR DRYING 8:48
CYCLE COMPLETE 8:51

Digital Signature #
212BEEB1DC0379CC
