
Midmark M11 - v1.0.5
Total Cycles: 622

Name: RITTER2
Sterilizer ID: V2313228
12/09/2021 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.1 F 0.0
2:00 151.5 F 2.1
4:00 167.2 F 6.3
6:00 189.2 F 11.3
8:00 215.1 F 5.0
10:00 228.9 F 10.1
12:00 233.1 F 8.3
14:00 243.6 F 13.3
16:00 243.7 F 12.5
18:00 256.8 F 19.4
20:00 268.4 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.2 F 29.3
2:00 271.7 F 29.5
3:00 272.0 F 29.2
4:00 271.6 F 29.4

Min 271.2 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 226.1 F 0.0
10:00 233.7 F 0.0
15:00 224.1 F 0.0
20:00 221.2 F 0.0
25:00 224.7 F 0.0
30:00 218.4 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 21:10
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

12/09/2021 03:20 PM
ACCEPTED: X
REJECTED:
Operator: U02