
Midmark M11 - v1.0.5
Total Cycles: 2358

Name: RITTER2
Sterilizer ID: V2313228
05/12/2023 07:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.0 F 0.0
2:00 80.6 F 0.0
4:00 120.1 F 0.0
6:00 123.8 F 0.3
8:00 157.0 F 0.4
10:00 178.3 F 0.4
12:00 194.6 F 3.3
14:00 216.6 F 6.5
16:00 226.2 F 10.0
18:00 234.9 F 10.4
20:00 234.2 F 9.6
22:00 239.7 F 10.5
24:00 243.1 F 12.2
26:00 240.7 F 10.8
28:00 252.7 F 16.9
30:00 263.2 F 23.2
32:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.4
2:00 271.5 F 28.7
3:00 272.5 F 29.3
4:00 271.5 F 28.6

Min 271.4 F 28.5
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 201.0 F 0.0
10:00 217.5 F 0.0
15:00 221.0 F 0.0
20:00 230.7 F 0.0
25:00 228.5 F 0.0
30:00 228.2 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:20
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:10:18
CYCLE COMPLETE

05/12/2023 09:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____