
Midmark M11 - v1.0.5
Total Cycles: 2039

Name: RITTER2
Sterilizer ID: V2313228
02/13/2023 10:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.0 F 0.0
2:00 165.0 F 0.1
4:00 176.0 F 1.1
6:00 199.6 F 5.4
8:00 216.8 F 10.4
10:00 225.9 F 6.7
12:00 238.1 F 12.0
14:00 237.5 F 10.1
16:00 237.0 F 8.7
18:00 250.8 F 16.1
20:00 262.5 F 22.9
22:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.5 F 29.5
2:00 271.5 F 28.9
3:00 272.7 F 29.6
4:00 271.8 F 29.0

Min 271.4 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 225.8 F 0.0
10:00 206.6 F 0.0
15:00 208.4 F 0.0
20:00 219.9 F 0.0
25:00 223.5 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 22:19
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:27
CYCLE COMPLETE

02/13/2023 11:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____