

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3417

Name: RITTER2  
Sterilizer ID: V2313228  
02/28/2024 10:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.8 F 0.0  
2:00 69.0 F 0.0  
4:00 71.1 F 0.0  
6:00 85.2 F 0.4  
8:00 121.5 F 2.0  
10:00 150.1 F 4.9  
12:00 171.9 F 7.9  
14:00 197.5 F 11.1  
16:00 215.1 F 5.8  
18:00 217.4 F 7.5  
20:00 227.7 F 11.4  
22:00 228.0 F 7.5  
24:00 240.8 F 12.8  
26:00 239.6 F 11.2  
28:00 251.3 F 16.9  
30:00 261.5 F 22.9  
32:00 271.2 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.5  
1:00 272.5 F 30.4  
2:00 272.4 F 29.9  
3:00 271.2 F 29.4  
4:00 272.1 F 30.4  
  
Min 271.2 F 29.4  
Max 273.1 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.3 F 0.0  
5:00 210.6 F 0.0  
10:00 201.7 F 0.0  
15:00 187.8 F 0.0  
20:00 185.9 F 0.0  
25:00 192.9 F 0.0  
30:00 210.1 F 0.0  
  
Drying Complete

FILLING: 0:51  
HEATING: 32:45  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:11:13  
CYCLE COMPLETE

02/28/2024 11:14 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_