
Midmark M11 - v1.0.5
Total Cycles: 205

Name: RITTER2
Sterilizer ID: V2313228
07/14/2021 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.6 F 0.0
2:00 171.3 F 2.4
4:00 170.3 F 5.9
6:00 189.6 F 10.1
8:00 206.5 F 6.7
10:00 219.3 F 8.7
12:00 232.3 F 12.8
14:00 235.6 F 11.3
16:00 238.9 F 10.0
18:00 251.1 F 16.2
20:00 261.9 F 22.4
22:00 271.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 271.5 F 29.8
2:00 273.0 F 30.1
3:00 271.1 F 29.5
4:00 272.4 F 30.4

Min 270.4 F 28.9
Max 273.4 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 234.1 F 0.0
10:00 230.8 F 0.0
15:00 227.9 F 0.0
20:00 230.9 F 0.0
25:00 231.1 F 0.0
30:00 179.9 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:54
CYCLE COMPLETE

07/14/2021 09:54 AM
ACCEPTED: X
REJECTED:
Operator: U01