

-----  
Midmark M11 - v1.0.5  
Total Cycles: 617

Name: RITTER2  
Sterilizer ID: V2313228  
12/09/2021 07:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.5 F 0.0  
2:00 74.9 F 0.0  
4:00 83.0 F 0.1  
6:00 107.2 F 1.3  
8:00 141.8 F 4.6  
10:00 164.8 F 8.7  
12:00 187.9 F 12.7  
14:00 219.8 F 4.6  
16:00 228.3 F 8.3  
18:00 240.5 F 12.6  
20:00 234.2 F 9.5  
22:00 239.4 F 9.8  
24:00 246.6 F 14.0  
26:00 258.2 F 20.3  
28:00 267.8 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.7 F 29.3  
1:00 273.0 F 30.0  
2:00 271.7 F 29.2  
3:00 272.9 F 29.8  
4:00 271.7 F 29.2  
  
Min 271.6 F 29.1  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.8 F 0.0  
5:00 223.6 F 0.0  
10:00 250.1 F 0.0  
15:00 224.7 F 0.0  
20:00 217.1 F 0.0  
25:00 238.7 F 0.0  
30:00 226.4 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 29:09  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 01:07:15  
CYCLE COMPLETE

12/09/2021 09:32 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02