
Midmark M11 - v1.0.0
Total Cycles: 1440

Name: RITTER 1
Sterilizer ID: V2178198
12/09/2021 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.9 F 0.0
2:00 180.4 F 3.1
4:00 190.8 F 7.0
6:00 215.4 F 13.0
8:00 224.4 F 7.0
10:00 236.9 F 12.0
12:00 235.2 F 8.9
14:00 239.6 F 9.9
16:00 247.1 F 13.7
18:00 259.7 F 20.6
20:00 271.1 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.6
2:00 272.0 F 28.6
3:00 272.1 F 28.5
4:00 271.9 F 28.6

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 220.4 F 0.0
10:00 226.2 F 0.0
15:00 223.2 F 0.0
20:00 219.2 F 0.0
25:00 216.6 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:29
CYCLE COMPLETE

12/09/2021 11:20 AM
ACCEPTED: X
REJECTED:
Operator: U02