
Midmark M11 - v1.0.0
Total Cycles: 1872

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2022 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.6 F 0.0
2:00 170.0 F 3.1
4:00 189.1 F 7.1
6:00 213.5 F 12.6
8:00 224.9 F 7.0
10:00 237.0 F 12.1
12:00 235.8 F 9.1
14:00 239.0 F 9.5
16:00 248.1 F 14.2
18:00 260.4 F 21.0
20:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.6
2:00 272.0 F 28.7
3:00 271.9 F 28.4
4:00 272.1 F 28.8

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 215.9 F 0.0
10:00 173.3 F 0.0
15:00 164.6 F 0.0
20:00 157.6 F 0.0
25:00 151.4 F 0.0
30:00 148.5 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:34
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

04/26/2022 02:07 PM
ACCEPTED: X
REJECTED:
Operator: U02