
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000867
8:21 25/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:23
STERILIZING 4:28
136.0 C 220kPa 4:28
136.1 C 221kPa 4:59
136.2 C 222kPa 5:29
136.4 C 221kPa 5:59
136.2 C 223kPa 6:29
136.4 C 223kPa 6:59
136.4 C 222kPa 7:29
136.4 C 223kPa 7:59
136.4 C 223kPa 7:59
VENTING 7:59
AIR DRYING 8:50
CYCLE COMPLETE 9:19

Digital Signature #
C101A8403CA960CE
