
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000062
9:24 03/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:26
STERILIZING 3:44
136.0 C 221kPa 3:44
136.3 C 225kPa 4:14
136.2 C 223kPa 4:44
136.4 C 225kPa 5:14
136.5 C 226kPa 5:44
136.1 C 222kPa 6:14
136.3 C 226kPa 6:44
136.3 C 224kPa 7:14
136.3 C 224kPa 7:14
VENTING 7:14
AIR DRYING 8:08
CYCLE COMPLETE 23:08

Digital Signature #
E2951DCB1A7B23C3
