
Midmark M11 - v1.0.0
Total Cycles: 3856

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2023 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.9 F 0.0
2:00 167.7 F 2.1
4:00 175.4 F 5.9
6:00 201.6 F 11.1
8:00 216.9 F 4.7
10:00 230.9 F 9.7
12:00 234.8 F 8.8
14:00 243.3 F 12.4
16:00 241.8 F 11.1
18:00 255.3 F 17.9
20:00 267.0 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 271.8 F 28.1
2:00 272.5 F 28.5
3:00 272.3 F 28.8
4:00 271.4 F 27.9

Min 271.3 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 209.7 F 0.0
10:00 214.3 F 0.0
15:00 205.8 F 0.0
20:00 201.1 F 0.0
25:00 208.5 F 0.0
30:00 210.7 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

10/03/2023 04:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____