
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007022
10:28 01/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:58
STERILIZING 2:58
136.4 C 220kPa 2:58
136.4 C 222kPa 3:28
136.4 C 223kPa 3:58
136.3 C 225kPa 4:28
136.3 C 223kPa 4:58
136.2 C 222kPa 5:28
136.4 C 222kPa 5:58
136.2 C 223kPa 6:28
136.2 C 223kPa 6:28
VENTING 6:28
AIR DRYING 7:19
CYCLE COMPLETE 7:27

Digital Signature #
D6EB020932BBA55E
