

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4679

Name: RITTER 1  
Sterilizer ID: V2178198  
06/26/2024 12:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 125.1 F 0.0  
2:00 128.3 F 1.8  
4:00 148.2 F 4.6  
6:00 173.3 F 7.7  
8:00 197.6 F 11.3  
10:00 210.7 F 3.6  
12:00 220.8 F 7.2  
14:00 231.2 F 11.1  
16:00 229.9 F 6.8  
18:00 240.5 F 12.0  
20:00 238.1 F 9.8  
22:00 250.0 F 15.6  
24:00 260.5 F 21.4  
26:00 270.2 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.4 F 29.4  
2:00 272.0 F 28.9  
3:00 271.2 F 28.6  
4:00 272.1 F 29.4

Min 271.2 F 28.5  
Max 272.4 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 218.3 F 0.0  
10:00 221.9 F 0.0  
15:00 216.9 F 0.0  
20:00 221.0 F 0.0  
25:00 226.6 F 0.0  
30:00 222.8 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 26:58  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:05:47  
CYCLE COMPLETE

06/26/2024 01:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_