
Midmark M11 - v1.0.0
Total Cycles: 5221

Name: RITTER 1
Sterilizer ID: V2178198
01/06/2025 12:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.2 F 0.0
2:00 104.4 F 0.0
4:00 122.8 F 0.0
6:00 153.5 F 0.3
8:00 180.1 F 0.6
10:00 199.3 F 3.3
12:00 208.8 F 6.6
14:00 218.0 F 10.0
16:00 227.5 F 10.7
18:00 231.3 F 8.5
20:00 241.5 F 13.0
22:00 239.3 F 10.3
24:00 237.7 F 8.9
26:00 248.7 F 14.3
28:00 259.3 F 20.1
30:00 268.5 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.1 F 28.6
2:00 272.7 F 28.9
3:00 272.3 F 28.5
4:00 271.9 F 28.2

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 206.2 F 0.0
10:00 203.9 F 0.0
15:00 205.2 F 0.0
20:00 201.5 F 0.0
25:00 200.9 F 0.0
30:00 201.5 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:27
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:10:09
CYCLE COMPLETE

01/06/2025 01:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____