
Midmark M11 - v1.0.0
Total Cycles: 5344

Name: RITTER 1
Sterilizer ID: V2178198
02/19/2025 03:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.4 F 0.0
2:00 126.7 F 1.0
4:00 133.5 F 3.6
6:00 156.1 F 6.4
8:00 172.5 F 10.1
10:00 195.1 F 8.0
12:00 213.6 F 6.2
14:00 222.7 F 10.2
16:00 230.8 F 8.3
18:00 235.2 F 10.3
20:00 236.6 F 8.8
22:00 245.2 F 13.4
24:00 256.2 F 19.2
26:00 266.3 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.3 F 29.7
2:00 273.0 F 29.7
3:00 272.2 F 29.2
4:00 271.4 F 29.0

Min 271.4 F 28.9
Max 273.4 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 183.2 F 0.0
10:00 202.2 F 0.0
15:00 188.8 F 0.0
20:00 205.3 F 0.0
25:00 220.2 F 0.0
30:00 225.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 27:48
STERILIZING: 4:00
VENTING: 3:60
DRYING: 30:00
TOTAL CYCLE: 01:06:39
CYCLE COMPLETE

02/19/2025 04:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____