

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1587

Name: RITTER 1  
Sterilizer ID: V2178198  
02/02/2022 09:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.6 F 0.0  
2:00 159.3 F 2.8  
4:00 176.2 F 6.8  
6:00 199.2 F 11.8  
8:00 215.7 F 5.1  
10:00 229.9 F 10.2  
12:00 233.5 F 7.6  
14:00 243.0 F 12.7  
16:00 241.6 F 11.0  
18:00 254.9 F 17.8  
20:00 266.4 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.7 F 29.0  
2:00 271.4 F 28.2  
3:00 272.6 F 29.1  
4:00 272.3 F 28.6  
  
Min 271.4 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
5:00 232.7 F 0.0  
10:00 224.4 F 0.0  
15:00 223.2 F 0.0  
20:00 227.1 F 0.0  
25:00 231.1 F 0.0  
30:00 229.4 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 21:37  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:55  
CYCLE COMPLETE

02/02/2022 10:19 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_