

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008324  
12:10 19/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:48  
PRESSURIZING 2:50  
STERILIZING 3:56  
136.4 C 221kPa 3:56  
136.4 C 224kPa 4:26  
136.2 C 223kPa 4:56  
136.4 C 224kPa 5:26  
136.4 C 224kPa 5:56  
136.4 C 224kPa 6:26  
136.4 C 223kPa 6:56  
136.1 C 222kPa 7:26  
136.1 C 222kPa 7:26  
VENTING 7:26  
AIR DRYING 8:17  
CYCLE COMPLETE 23:52

-----  
Digital Signature #  
8345A37C1517DCF6  
-----