
Midmark M11 - v1.0.5
Total Cycles: 1000

Name: RITTER2
Sterilizer ID: V2313228
04/18/2022 02:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.5 F 0.0
2:00 186.1 F 1.4
4:00 204.0 F 6.8
6:00 219.7 F 12.5
8:00 230.0 F 8.8
10:00 236.7 F 10.2
12:00 244.5 F 13.5
14:00 244.9 F 13.1
16:00 258.1 F 20.1
18:00 269.5 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.8
1:00 272.3 F 29.6
2:00 271.5 F 28.7
3:00 272.4 F 29.5
4:00 271.6 F 28.7

Min 271.3 F 28.7
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 222.7 F 0.0
10:00 226.3 F 0.0
15:00 230.3 F 0.0
20:00 227.9 F 0.0
25:00 228.9 F 0.0
30:00 228.1 F 0.0

Drying Complete

FILLING: 1:39
HEATING: 18:50
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:43
CYCLE COMPLETE

04/18/2022 03:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____