
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006308
15:40 15/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:08
STERILIZING 4:14
136.4 C 223kPa 4:14
136.4 C 224kPa 4:44
136.4 C 224kPa 5:14
136.1 C 221kPa 5:44
135.9 C 218kPa 6:14
136.4 C 224kPa 6:44
136.4 C 224kPa 7:14
136.4 C 222kPa 7:44
136.4 C 222kPa 7:44
VENTING 7:44
AIR DRYING 8:32
CYCLE COMPLETE 21:58

Digital Signature #
4F716469AF594D8E
