
Midmark M11 - v1.0.5
Total Cycles: 2615

Name: RITTER2
Sterilizer ID: V2313228
07/06/2023 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.0 F 0.0
2:00 156.2 F 0.0
4:00 155.3 F 0.2
6:00 184.0 F 3.5
8:00 213.1 F 8.1
10:00 219.7 F 13.1
12:00 232.2 F 8.3
14:00 243.1 F 13.4
16:00 241.0 F 11.3
18:00 238.5 F 9.9
20:00 251.7 F 16.4
22:00 262.8 F 22.9
24:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.5 F 29.5
2:00 271.7 F 28.7
3:00 271.8 F 29.2
4:00 272.7 F 29.3

Min 271.0 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 234.3 F 0.0
10:00 227.6 F 0.0
15:00 225.0 F 0.0
20:00 227.8 F 0.0
25:00 225.3 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 24:22
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:22
CYCLE COMPLETE

07/06/2023 10:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____